



Product Code . JL-OAP-1704

Flash Point Pensky Martens Closed

Description

Flash Point Pensky Matrens Closed

Flash Point Pensky Martens (Closed) apparatus with NPL Certificate, Electrical Heating, with gas test jet and electric heater with ZI P 10303.

Energy regulator.

Suitable for operation on 22V, single phase, 50 Hz, AC supply.

Complete in wooden carrying case.

Flash Point (Closed) Pensky-Martens Apparatus same as but without NPL Certificate.

Flash Point (Open) & Fire Point Pensky-Martens Apparatus with NPL Certificate.

Electrical heating, with gas test jet and electric heater with energy regulator.

Suitable for operation on 220, 50 Hz, Single Phase, AC Supply.

Complete in a wooden carrying case.

Essential Extras

Thermometer IP 15C Range: -5° C to +110°C

Thermometer IP 16C Range: 90°C to +370°C

Replacement Parts

Pensky-Martens Oil Cup

Pensky-Martens Lid shutter with Gas Test Jet

