



Product Code . JL-PE-3823

Shortwave Medical Diathermy

Description

Shortwave Medical Diathermy

The Filament Voltage can be adjusted in six steps.

It has 30min treatment timer.

The operating frequency 27.12 MHz.

Maximum output power 500Watts.

Switched off RF power at the end of the treatment.
