



**Product Code . JL-PE-3831**

## Diagnostic Muscle Stimulator

### Description

---

Diagnostic Muscle Stimulator

Input Voltage: 220V AC 50Hz

Maximum output 120 Volts, smooth intensity control

Pulse Width: 10 micro/sec. to 300 millisecc.

Pulse duration step control .01 to 300 millisecc.

Frequency: 6 Pulse per min. to 30 pulse per min. 1 pulse/millisecc. to 100 pulse millisecc.

Faradic: Comprises 40 pulses per second, each pulse of 0.9 ms

Surge Faradic: Two isolated controls are provided for changing contraction & rest period

Intensity: Common intensity control for all four mode of therapy..

Mode Selector: Logic selector for selection of each therapy mode.

An analog meter is provided for display of intensity for all four modes.

---