

## Jain Laboratory Instruments







**Product Code . JL-PE-3842** 

## Digital Ultrasonic Physiotherapy Equipment

## **Description**

Digital Ultrasonic Physiotherapy Equipment

Power Output:- 17 watt continuous, 22 watt pulsed, 29 watt pulsed.

**Operating Frequency:-** 1 Mhz.

Input Voltage:- 220/240 Volts A.C. 50Hz.