

Jain Laboratory Instruments







Product Code . JL-XRMAE-3902

Automatic X Ray Machine

Description

Automatic X Ray Machine

Radio graphic output: 24Kw (AMU-300) and 40kw (AMU-500) as per IS7620 (parts).

Tube filament current stabilization through constant voltage transformer (CVT).

Full wave rectification.

Digital display of KV, mA and MAS.

Overload protection (blockage of exposure).

Shortest exposure time 10 MAS in 500mA.

Supports serialography up to 6fps kv, mA and MAS values set with step switches fluoroscopy time totalize.

5 LEO indicators for Line, 0/L exposure, ready and bucky in voltage stabilization (compensation).

Through course and fine voltage compensation switches voltmeter and mA meter.